

DECLARATION OF CONFORMITY

MANUFACTURER: ALCOHOL COUNTERMEASURE SYSTEMS CORP.
ADDRESS: 60 International Boulevard, Toronto, Ontario M9W 6J2, CANADA
PRODUCT: Passive Alcohol Tester
MODEL: ALCOSCAN™
INTENDED USE: Passive alcohol testing
POWER SUPPLY: 4 X 1.5 volt (AA alkaline batteries)
CLASSIFICATION: Group 1, Class B

The device noted above has been tested with the essential requirements and other relevant provisions of EMC Directive 2004/108/EC when used for its intended purpose:

Electromagnetic compatibility pursuant to: EN 55011:2007 + A2	Applied standard(s): EN 55011:2007 + A2 (Emissions) EN 61326-1: 2006 (ESD Susceptibility)
The device also meets the requirements for IP 52 according to IEC 60529, except for the breath sampling port. Alcohol Countermeasure Systems Corp. (ACS) hereby declares that the ALCOSCAN complies with above Directives. The following marking applies to the above mentioned product.	

PLACE OF ISSUE: Toronto, Ontario, Canada
DATE OF ISSUE: 2015.02.04
ISSUED BY: Abe Verghis, Regulatory Affairs Manager


SIGNATURE:



NOTARIAL CERTIFICATE OF EXECUTION

I, STEVEN H. LEACH, a Notary Public for the Province of Ontario, do certify that on this day, before me personally appeared Abe Verghis, to me know and known to me to be the individual described in the attached instrument, who, in my presence, executed the attached instrument and who, by me being sworn, did acknowledge that he executed this instrument as his free act and deed.

IN TESTIMONY WHEREOF I have hereto subscribed my name and affixed my seal of Office at Toronto, Ontario, Canada, this 04th day of February, 2015.



Steven H. Leach

ALCOSCAN is a trademark of Alcohol Countermeasure Systems (International) Inc. and used under license.